

GASTEC Instructions for No.15L Nitric Acid Detector Tube

FOR SAFE OPERATION :

Read this manual and the instruction manual of your Gastec Gas Sampling Pump carefully.

⚠ WARNING :

1. Use only Gastec detector tubes in a Gastec Pump.
2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
3. The use of non-Gastec parts or components in Gastec's detector tube and pump system or use of a non-Gastec detector tube with a Gastec pump or use of a Gastec detector tube with a non-Gastec pump may result in property damage, serious bodily injury, and death; voids all warranties; and voids all performance and data accuracy guaranties.

⚠ CAUTION : If not observed, injuries to the operator or damage to the product may result.


1. When breaking the tube ends, keep away from eyes.
2. Do not touch the broken glass tubes, pieces and reagent with bare hand(s).
3. The sampling time represents the time necessary to draw the air sample through the tube.
The tube must be positioned in the desired sampling area for the entire sampling time or until the flow finish indicator indicates the end of the sample.

⚠ NOTES : For maintaining performance and reliability of the test results.

1. Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
2. Use this tube within the temperature range of 0 - 40°C (32 - 104°F).
3. Use this tube within the absolute humidity range of 4 - 16mg/L.
4. This tube may be interfered with by the coexisting gases. Please refer to the "INTERFERENCES".
5. Shelf life and storage conditions of the tube are marked on the label of the box of tube.

APPLICATION OF THE TUBE : Use of this tube for the detection of Nitric Acid in air or the industrial areas and environmental atmospheric condition.

SPECIFICATION : (As a result of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	0.1 - 1 ppm	1 - 20 ppm	20 - 40 ppm
Number of Pump Strokes	2 - 10	1	1/2
Correction Factor	1/2 - 1/10	1	2
Sampling Time	45 seconds per pump stroke		30 seconds
Detecting Limit	0.05 ppm (n = 10)		
Colour Change	Yellow → Reddish purple		
Reaction Principle	HNO ₃ + Indicator → Reaction Products		

Coefficient of Variation : 15% (for 1 to 5 ppm), 10% (for 5 to 20 ppm)

**** Shelf Life : Please refer to the Validity Date printed on the box of tube.**

**** Store the tubes in dark and cool place.**

CORRECTION FOR TEMPERATURE, HUMIDITY AND PRESSURE :

Temperature : No correction is required.

Humidity : Correct for humidity by the table below:

Absolute Humidity (mg/L)	4	6	8	10	12	14	16
Correction Factor	0.55	0.68	0.85	1.0	1.17	1.3	1.35

Pressure : To correct for pressure, multiply the tube reading by

$$\frac{\text{Tube Reading (ppm)} \times 1013 \text{ (hPa)}}{\text{Atmospheric Pressure (hPa)}}$$

MEASUREMENT PROCEDURE :

1. For leak checking of the pump insert a fresh sealed detector tube into pump. Follow instructions provided with the pump operating manual.
2. Break tips off a fresh detector tube in the tube tip breaker of the pump.
3. Insert the tube into the pump inlet with arrow on the tube pointing toward pump.
4. Make certain pump handle is all the way in. Align guide marks on pump body and handle.
5. Pull the handle all the way out until it locks on 1 pump stroke (100ml). Wait 45 seconds and confirm the completion of the sampling.
6. For lower than 1 ppm measurement, repeat the above sampling procedure 1 - 9 more time until the stain attains to the first calibration mark. For higher than 20 ppm measurement, prepare fresh tube and take 1/2 pump stroke.
7. Read concentration at the interface of the stained-to-unstained reagent.
8. If correction is needed, multiply the correction factors of humidity, pump strokes and pressure.

INTERFERENCES :

Substance	Concentration	Interference	Changes colour by itself to
Hydrogen chloride	≥ 1/10	+	Reddish purple
Chlorine		No	Bleach
Carbon monoxide		No	No
Nitrogen dioxide		No	No
Organic solvents		No	No

The table of this interference gases primarily expresses the interference of each coexisting gas in the gas concentration range, equivalent to the gas concentration. Therefore, the test result may be given positive result by the other substances not listed in the table. For more information is needed, contact us or our distributors in your territory.

APPLICATION FOR OTHER SUBSTANCES:

Tube 15L can also be used for other substances as below:

Substance	Correction Factor	No. of pump strokes	Measuring range
Hydrogen bromide	0.8	1	0.8 to 16 ppm
Trichloroacetic acid	Refer to scale below	1	1 to 37.5 ppm
Trichloroacetic Acid Conc. (ppm)			
Tube No.15L Reading (n=1)			

CORRECTION FACTOR:

Detector tubes are primarily designed to measure specific gases. But it is also possible to measure other substances of similar chemical properties with the aid of a correction factor or chart. Therefore, please make use of the correction factor/chart measuring ranges as a reference. For more precise factor please contact your Gastec distributor.

DANGEROUS AND HAZARDOUS PROPERTIES:

Threshold Limit Value-Time Weighted Average by ACGIH (2006) : 2 ppm
 Threshold Limit Value-Short Term Exposure Limit: by ACGIH (2006) : 4 ppm

DISPOSAL INSTRUCTION:

Reagent of the tube does not use toxic substances. When dispose of the tube regardless of whether used or unused, follow the rules and regulations of the local government.

WARRANTY:

If you have any questions regarding gas detection and quality of the tubes, please feel free to contact your Gastec representatives.

Manufacturer : Gastec Corporation
 8-8-6 Fukayanaka, Ayase-City, 252-1195, Japan
<http://www.gastec.co.jp/>
 Telephone +81-467-79-3910 Facsimile +81-467-79-3979

IM0015LE1
 Printed in Japan
 07F1Z