

Performance

| Measuring range | 10 to 400 ppm | 400 to 1120 ppm |
|------------------------|---------------|-----------------|
| Number of pump strokes | 2(200 mL) | 1(100 mL) |
| Correction factor | 1 | 2.8 |
| Sampling time | 4 min | 2 min |

 $\begin{array}{lll} \mbox{Detecting limit:} & 2 \mbox{ ppm } (2 \mbox{ pump strokes}) \\ \mbox{Colour change:} & \mbox{Yellow} \rightarrow \mbox{Pale blue} \end{array}$

Operating conditions : Temperature 0 to 40 °C (32 to 104 °F) correction used Relative humidity 0 to 90 % correction not used

Relative standard deviation: 10 % (for 10 to 100 ppm), 5 % (for 100 to 400 ppm)

Tube quantity and number of tests per box: 10 tubes for 10 tests

Shelf life: 24 months

Reaction principle

 $(C_2H_5)_2O + Cr^6 + H_2SO_4 \rightarrow Cr^3 +$

Possible coexisting substances and their interferences

| Substance | Concentration | Interference | Changes colour by itself to |
|---------------|---------------|--------------|-----------------------------|
| Methanol | ≥ 10 ppm | + | Pale blue (≥ 10 ppm) |
| Acetone | ≥ 10 ppm | + | Pale blue (≥ 10 ppm) |
| Ethyl acetate | ≥ 1.6 ppm | + | Pale blue (≧ 1.6 ppm) |

Calibration gas generation

Diffusion tube method