


Chloride Ion Cl^-

No.221L



Performance The minimum scale value (25mg/l) is not printed on the tube, but only the scale line is printed.

Measuring range	25 to 1000 mg/l
Sampling time	3 minutes
Detecting limit :	1 mg/l
Colour change :	Brown → White
Corrections for water temperature :	Unnecessary (0 to 50°C)
pH value :	pH 3.0 to pH 11.0
Relative standard deviation :	15 % (for 25 to 300 mg/l), 10 % (for 300 to 1000 mg/l)
Shelf life :	3 years

Reaction principle



Possible coexisting substances and their interferences

Substance	Concentration	Interference	Changes colour by itself to
Cyanide ion CN^-	≥ 1 mg/l	+	White
Bromide ion Br^-	≥ 10 mg/l	+	White
Thiocyanate ion SCN^-	≥ 30 mg/l	+	White
Iron(II) Fe^{2+}	≥ 200 mg/l	-	No (≤ 1000 mg/l)
Iron(III) Fe^{3+}	≥ 200 mg/l	-	White
Fluoride ion F^-	≥ 500 mg/l	-	No (≤ 1000 mg/l)
Iodide ion I^-	≥ 50 mg/l	+	White
Sulphide ion S^{2-}	≥ 5 mg/l	+	Blackish gray
Sulphate ion SO_4^{2-}	≥ 2000 mg/l	-	White

Other substance measurable with this detector tube

Substance	Correction	Measuring range
Bromide ion Br^-	Factor : 2.2	55 to 2200 mg/l

Calibration method

Sodium chloride ion standard solution