

## Detector tube No. 102L, Hexane, Instruction Manual Change

Gastec Corporation

Number of pump strokes and correction factor for will change, and also associated measuring range will change.

●Details of the change to the Instruction Manual:

Before the change:

Measuring Range	4 – 50 ppm	50 – 1200 ppm
Number of Pump Strokes	5	1
Correction Factor	1/12	1

After the change:

Measuring Range	3.5 – 9 ppm <sup>*1</sup>	9 – 25 ppm	25 – 50 ppm	50 – 1200 ppm
Number of Pump Strokes	8	4	2	1
Correction Factor	0.07	0.18	0.5	1

\*1) If this tube is used with 8 pump strokes, it should be used in an environment with an absolute humidity of 7 mg/L (at 20-25°C, RH 30%) or less.

●Contact Information

GASTEC CORPORATION

8-8-6 Fukayanaka, Ayase-City, Kanagawa, 252-1195, JAPAN

Tel: +81(0)467-79-3910 Fax: +81(0)467-79-3979