


Dosi-tube


Tetrachloroethylene Cl₂C:CCl₂ No.133D



Performance

Measuring range	3 to 150 ppm
Sampling time	1 to 8 hours
Colour change :	Yellow → Purple
Operating conditions :	Temperature 0 to 40 °C (32 to 104 °F) correction used Relative humidity 20 to 80 % correction not used
Relative standard deviation :	15 % (for 25 to 50 ppm·hr), 10 % (for 50 to 150 ppm·hr)
Tube quantity and number of tests per box :	10 tubes for 10 tests
Shelf life :	15 months (in the refrigerator)

Reaction principle



Possible coexisting substances and their interferences

Substance	Concentration	Interference	Changes colour by itself to
Hydrogen chloride		+	Purple
Chlorine		+	
1,2-Dichloroethylene		+	
Trichloroethylene		+	
Toluene, Xylene		No	No

Calibration gas generation

Diffusion tube method