

# GASTEC Instructions for No.12M Hydrogen Cyanide Detector Tube

## FOR SAFE OPERATION :

Read this manual and the instruction manual of your Gastec Gas Sampling Pump carefully.

### ⚠ WARNING :

1. Use only Gastec detector tubes in a Gastec Pump.
2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
3. The use of non-Gastec parts or components in Gastec's detector tube and pump system or use of a non-Gastec detector tube with a Gastec pump or use of a Gastec detector tube with a non-Gastec pump may result in property damage, serious bodily injury, and death; voids all warranties; and voids all performance and data accuracy guaranties.

### ⚠ CAUTION : If not observed, injuries to the operator or damage to the product may result.

1. When breaking the tube ends, keep away from eyes.
2. Do not touch the broken glass tubes, pieces and reagent with bare hand(s).
3. The sampling time represents the time necessary to draw the air sample through the tube. The tube must be positioned in the desired sampling area for the entire sampling time or until the flow finish indicator indicates the end of the sample.

### △ NOTES : For maintaining performance and reliability of the test results.


1. Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
2. Use this tube within the temperature range of 0 - 40°C (32 - 104°F).
3. Use this tube within the relative humidity range of 0 - 90%.
4. This tube may be interfered with by the coexisting gases. Please refer to the "INTERFERENCES".
5. Shelf life and storage conditions of the tube are marked on the label of the box of tube.

## APPLICATION OF THE TUBE :

Use of this tube for the detection of Hydrogen cyanide in air or the industrial areas and environmental atmospheric condition.

## SPECIFICATION :

(As a result of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	17 - 50 ppm	50 - 800 ppm	800 - 2400 ppm
Number of Pump Strokes	2	1	1/2
Correction Factor	1/3	1	3
Sampling Time	1 minute per pump stroke		30 seconds
Detecting Limit	1 ppm ( n = 2 )		
Color Change	Yellow → Red		
Reaction Principle	Hydrogen cyanide reacts with mercuric chloride to form the hydrogen chloride then discolors the indicator to red. $2HCN + HgCl_2 \rightarrow Hg(CN)_2 + 2HCl$ HCl + Base → Chloride product		

**Coefficient of Variation:** 10% (for 50 to 200 ppm), 5 % (for 200 to 800 ppm)

**\*\* Shelf Life :** Please refer to the Validity Date printed on the box of tube.

**\*\* Store the tubes in dark and cool place.**

## CORRECTION FOR TEMPERATURE, HUMIDITY AND PRESSURE :

**Temperature :** No correction is required.

**Humidity :** No correction is required.

**Pressure :** To correct for pressure, multiply the tube reading by  

$$\frac{\text{Tube Reading (ppm)} \times 1013 \text{ (hPa)}}{\text{Atmospheric Pressure (hPa)}}$$

## MEASUREMENT PROCEDURE :

1. For leak checking of the pump insert a fresh sealed detector tube into pump. Follow instructions provided with the pump operating manual.
2. Break tips off a fresh detector tube in the tube tip breaker of the pump.
3. Insert the tube into pump inlet with arrow on the tube pointing toward pump.
4. Make certain pump handle is all the way in. Align guide marks on pump body and handle.
5. Pull the handle all the way out until it locks on 1 pump stroke (100ml). Wait 1 minute and confirm the completion of the sampling.
6. For lower than 50 ppm measurement, repeat the above sampling procedure one more times until the stain attains to the first calibration marks. For higher than 800 ppm measurement, prepare fresh tube, and take 1/2 pump stroke.
7. Read concentration at the interface of the stained-to-unstained reagent.
8. If correction is needed, multiply the correction factors of pump strokes and pressure.

## INTERFERENCES :

Substance	Concentration	Interference	Changes color by itself to
Hydrogen sulfide	≥ 500 ppm	+	Red
Sulfur dioxide	≥ 500 ppm	+	Red

The table of this interference gases primarily expresses the interference of each coexisting gas in the gas concentration range, equivalent to the gas concentration. Therefore, the test result may be given positive result by the other substances not listed in the table. For more information is needed, please contact us or our distributors in your territory.

## DANGEROUS AND HAZARDOUS PROPERTIES :

Threshold Limit Value-Ceiling by ACGIH (2005) : 4.7 ppm.

Explosive Range : 5.6 - 40%

## DISPOSAL INSTRUCTION :

Reagent of the tube uses toxic inorganic mercury as reagent. When dispose of the tube regardless of whether used or unused, follow the rules and regulations of the local government.

## WARRANTY :

If you have any questions regarding gas detection and quality of the tubes, please feel free to contact your Gastec representatives.

Manufacturer : Gastec Corporation  
 8-8-6 Fukayanaka, Ayase-City, 252-1195, Japan  
<http://www.gastec.co.jp/>  
 Telephone +81-467-79-3910 Facsimile +81-467-79-3979

IM0012ME1  
 Printed in Japan  
 05L1Z