

Detector tube No. 60, Phenol, Instruction Manual Correction

Gastec Corporation

We would like to apologize for the error discovered in the Detector Tube, No. 60, Phenol, Instruction Manual regarding the Correction Factor.

We have corrected the error in the Correction Factor, and also changed the associated Measuring Range.

●Details of the correction to the Instruction Manual:

Before the correction:

Measuring Range	0.4 – 1 ppm	1 – 25 ppm	25 – 62.5 ppm	62.5 – 187 ppm
Number of Pump Stroke (n)	4	2	1	1/2
Correction Factor	0.4	1	2.5	7.5

*The highlighted box is the Scale Range.

Tube Reading (ppm)	True Concentration (ppm)						
	10°C (50°F)	15°C (59°F)	20°C (68°F)	25°C (77°F)	30°C (86°F)	35°C (95°F)	40°C (104°F)
25	38	30	25	23.5	23	22.5	21
20	28	23	20	19	19	18.5	17
15	18	17	15	14.5	14	14	13
10	11	11	10	10	10	10	8
5	5	5	5	5	5	5	5
3	3	3	3	3	3	3	3
1	1	1	1	1	1	1	1

After the correction:

Measuring Range	0.12 – 1 ppm	1 – 25 ppm	25 – 70 ppm	70 – 183 ppm
Number of Pump Stroke (n)	4	2	1	1/2
Correction Factor	0.12	1	2.8	7.3

*The highlighted box is the Scale Range.

Tube Reading (ppm)	True Concentration (ppm)						
	10°C (50°F)	15°C (59°F)	20°C (68°F)	25°C (77°F)	30°C (86°F)	35°C (95°F)	40°C (104°F)
25	35	29.5	25	23.5	21.8	21.6	20.4
20	28	23.6	20	18.8	17	16	15
15	21	17.7	15	14	13	11	10
10	14	12	10	9.4	8.7	6.7	6.2
5	7	6	5	4.7	4.4	3	2.8
3	4.2	3.5	3	2.8	2.6	1.8	1.6
1	1.4	1.2	1	0.94	0.87	0.6	0.5

We apologize for this error and thank you for your understanding.